

Superspeed Rotors

# Thermo Scientific TCF-20 Continuous Flow and Zonal Rotors



## ROTOR SPECIFICATIONS

<b>Type</b>	Continuous Flow / Zonal	
<b>Material</b>	Titanium	
<b>Capacity</b>	1350 mL	
<b>Maximum Speed</b>	20,000 rpm	
<b>K Factor</b>	620	
<b>Net Weight</b>	7.8 kg/17.2 lbs	
<b>Warranty</b>	5 years	
	<b>RCF</b>	<b>RADIUS</b>
<b>Maximum</b>	42,931 x g	96.0 mm
<b>Minimum</b>	16,099 x g	36.0 mm

## ROTOR AND ACCESSORIES ORDERING INFORMATION

		THERMO SCIENTIFIC SORVALL LYNX 6000
<b>TCF-20 Rotor</b>	Max. Speed	20,000 rpm
	Max. RCF	42,931 x g

### TCF-20 Continuous Flow Rotor Package includes:

Description	Cat. No.
TCF-20 Continuous Flow Rotor Package	75003012
Titanium Rotor Body	75000604
Rotor Lid	75004012
Septa	49609
Tube Holder Assembly	75000601
Tool Kit	75004052
Continuous Flow Assembly	75000605
Instruction Manual	50139705

### TCF-20 Zonal Flow Rotor Package includes:

Description	Cat. No.
TCF-20 Zonal Flow Rotor Package	75003013
Titanium Rotor Body	75000604
Rotor Lid	75004012
Septa	49609
Tube Holder Assembly	75000601
Tool Kit	75004052
Zonal Mode Assembly	75004051
Instruction Manual	50139705

## ACCESSORIES ORDERING INFORMATION

### Accessories:

Description	Cat. No.
Peristaltic Pump	20210448
Latex Tubing	13117
O-Ring Set for Continuous Flow Assembly	75004006
O-Ring Set for Zonal Mode Assembly	75004007
Screws and Small Parts	75004008
Bearing Kit	75004050
Tube Holder Assembly	75000601
Titanium Rotor Body	75000604
Rotor Lid	75004012
Septa	49609
Tool Kit	75004052
Continuous Flow Assembly	75000605
Zonal Mode Assembly	75004051
Rotor Unlocking Tool	75000602

Find out more at [thermofisher.com/rotors](http://thermofisher.com/rotors)

© 2013, 2014 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.

DSCFGLYNXTCF20 0617

**ThermoFisher**  
S C I E N T I F I C